



**International Journal of Biology, Pharmacy
and Allied Sciences (IJBPAS)**
'A Bridge Between Laboratory and Reader'

www.ijbpas.com

**DEVELOPMENT AND VALIDATION OF SIMPLE UV
SPECTROPHOTOMETRIC METHOD FOR ANALYZING
CIPROFLOXACIN HYDROCHLORIDE**

CHOUDHARY G^{1*} AND VERMA V²

1: Research Scholar, Kharvel Subharti College of Pharmacy, Swami Vivekanand Subharti
University, Meerut, 250005, India

2: Associate Professor, Kharvel Subharti College of Pharmacy, Swami Vivekanand Subharti
University, Meerut, 250005, India

***Corresponding Author: Goldy Choudhary: E Mail: goldy12345.gc@gmail.com**

Received 15th June 2020; Revised 20th July 2020; Accepted 23rd Oct. 2021; Available online 1st July 2022

<https://doi.org/10.31032/IJBPAS/2022/11.7.6206>

ABSTRACT

A simple, precise, accurate and rapid UV Spectroscopic method has been developed for the assay of Ciprofloxacin hydrochloride (HCL) in its pharmaceutical tablet dosage form. The UV Spectroscopic method was validated according to the International Conference on Harmonization (ICH) guideline, where range, linearity, accuracy, recovery, precision and sensitivity of the method were examined. Ciprofloxacin HCL showed maximum wavelength of absorbance at 272 nm in water and obeyed linearity or Beer-Lambert's law within the concentration range of 2-12 µg/mL. The regression of coefficient (R^2) was found to be greater than 0.99. The limit of detection (LOD) and limit of quantitation (LOQ) values were found to be 0.1µg/mL and 21µg/mL, respectively. The developed method was applied successfully for the analysis of ciprofloxacin HCl with good accuracy and precision.

Keywords: International Conference on Harmonization, Ciprofloxacin, Validation, LOD, LOQ

INTRODUCTION

The introduction of generic drug product from multiple sources into the health care delivery system of many developing countries was aimed at improving the overall healthcare delivery systems in such countries. The need to select one

product from among several generic drug products of the same active ingredients during the course of therapy is a cause of concern to a healthcare practitioner. The first stage in ascertaining the therapeutic equivalence of any drug product involves ascertaining the chemical and biopharmaceutical equivalency of such drug products [1].

Drug products that are chemically and biopharmaceutically equivalent must be identical in strength, quality, purity as well as content uniformity, disintegration and dissolution rates. Variable clinical response to the same dosage form of a drug product supplied by different manufacturers has been reported (Remington's Pharmaceutical Sciences, 1990). Therapeutic in-equivalences have been reported from the use of some generic brands of drug products such as tolbutamide [2].

There are several brands of Ciprofloxacin hydrochloride tablets available within the drug delivery system globally as well as in India. The increasing level of use of Ciprofloxacin hydrochloride tablets as a result of its versatility in the management of various cases of microbiological infections necessitated the need to evaluate the quality of the Ciprofloxacin hydrochloride tablets available in India. Ascertaining the quality of drug products involves the use of various procedures which includes

both biopharmaceutical and chemical assay techniques. Various methods have been reported for the chemical assay of Ciprofloxacin tablets [3].

In our study we used UV spectroscopy for assaying Ciprofloxacin HCl. This study is useful because this drug is commonly administered for various kinds of infections. The UV spectrophotometric analysis is often preferred in quality control testing and ordinary laboratories due to its broader availability, suitability, cost effectiveness and ease of use [4]. The aim of the present investigation is to develop a simple, sensitive and reproducible UV Spectrophotometric method for analysis of Ciprofloxacin HCl in a tablet dosage form and hence an economical method was developed and validated according to the ICH guidelines.

Ciprofloxacin (sip" roe flox' a sin) is an oral fluoroquinolone that is used to treat mild to-moderate urinary and respiratory tract infections. Ciprofloxacin is also used for infectious diarrhea, typhoid fever, uncomplicated gonorrhea, treatment of *Neisseria meningitidis* nasal carriage and prophylaxis against anthrax (RH, 2013). Like other fluoroquinolones, Ciprofloxacin is active against a wide range of aerobic gram-positive and gram-negative organisms. The fluoroquinolones are believed to act by inhibition of type II DNA topoisomerases (gyrases) that are required for synthesis of bacterial mRNAs (transcription) and

DNA replication. Ciprofloxacin was approved for use in the United States in 1990 and, currently, approximately 20 million prescriptions are filled yearly [5]. Ciprofloxacin is available in multiple oral formulations. Intravenous formulation. Common side effects include gastrointestinal upset, headaches, skin rash and allergic reactions. Less common, but more severe side effects include prolongation of the QT interval, seizures, hallucinations, tendon rupture, angioedema, Stevens Johnson syndrome and photosensitivity [6].

MATERIALS AND METHOD

2.1 Materials

2.1.1 Sample Collection

For the purpose of analyzing Ciprofloxacin HCL, reference standard was collected from the Eskaycf India Limited and tablets of Ciprofloxacin HCL of two different strengths (i.e., 250 mg ciprocin® and 500 mg ciprocin® tablet of Square Pharmaceutical Limited) from a reputed pharmaceutical company were purchased from retail pharmacy situated in, India (Table 1).

2.1.2 Sample

For conducting the UV method development study, the following solvents were used (Table 2).

2.1.3 Equipment's & Instruments

The equipment's and instruments used for the study were included in Table 3.

Table 1: Materials used for the experiment

Sample name	Source(Supplier Name)
Ciprofloxacin Standard	Eskayef India Limited
Ciprocin® 250mg	Square Pharmaceutical Limited
Ciprocin® 500mg	Square Pharmaceutical Limited

Table 2: Solvents used for developing UV method for Ciprofloxacin HCL

Reagents Name	Source (Supplier Name)
Methanol (Laboratory grade)	Active Fine Chemicals
Distilled Water/De-ionized water	Active Fine Chemicals

Table 3: Lists of equipment's used for the experiment

Equipment	Source (supplier name)	Origin
UV Spectrophotometer	Shimadzu UV1800	Japan
Electronic Balance	Shimadzu ATX 224	Japan
Sonicator	HWASHIN Power sonic 520	Korea

2.2 Methods

2.11 Wavelength Selection

Ciprofloxacin 100 µg/ml stock solution was accurately prepared by dissolving 5 mg of

Ciprofloxacin reference standard in 50 ml distilled water. A solution of 100 µg/ml concentration was prepared by diluting the

stock solution and scanned in the UV regions

(200-400 nm). The maximum wavelength of absorbance (λ_{max}) was observed at 272 nm with the absorbance value of 0.658. Therefore, 272 nm was selected as the λ_{max} for Ciprofloxacin HCL. However, in different papers the λ_{max} of Ciprofloxacin HCL were found to be slightly different [4].

2.2.2 Preparation of standard solution of Ciprofloxacin in water

Ciprofloxacin is freely soluble in water (Solubility is 10-30 mg/ml) at acidic pH (<5.0) [7]. Therefore, standard solution of Ciprofloxacin was prepared in water. A stock solution of 100 μ g/mL was prepared by accurately weighing 5 mg of Ciprofloxacin powder in to a 50 mL volumetric flask. The drug was then dissolved in distilled water. Required amount of water was added to make it volume. The stock solution was diluted further with distilled water to prepare concentrations of 2, 3, 4, 6, 8, 10, 12 μ g/ml [no. of replicates (n) were 3 for all cases].

2.2.3 Construction of calibration curve

Using UV Spectrophotometer, the absorbance values of these diluted solutions were noted at 272nm wavelength. All the samples were prepared and measured in three replicates. Data then plotted using MS Excel 2007 software. A calibration graph

was prepared by plotting the concentration versus the average absorbance values.

2.2.4 Method validation of Ciprofloxacin Hydrochloride

Linearity

"The linearity of an analytical procedure is its ability to elicit test results that are directly or by a well-defined mathematical transformation, proportional to the concentration of analyte in samples within a given range. Thus, in this section, "linearity" refers to the linearity of the relationship of concentration and assay measurement. In some cases, to attain linearity, the concentration and/or the measurement may be transformed".

The linearity was evaluated by analyzing three different calibration curves of Ciprofloxacin HCl. The regression coefficient and slope were specifically evaluated by using the RSD value.

2.2.4.2 Range

"The range of an analytical procedure is the interval between the upper and lower levels of analyte (including these levels) that have been demonstrated to be determined with a suitable level of precision, accuracy, and linearity using the procedure as written. The range is normally expressed in the same units as test results (e.g., percent, parts per million) obtained by the analytical procedure" [8].

The concentration range between which Ciprofloxacin HCl obeyed Beer-Lambert's law was taken as the range.

2.2.4.3 Accuracy testing

Accuracy testing was performed according to USP. Accuracy is calculated as the percentage of recovery by the assay of the known added amount of analyte in the sample, or as the difference between the mean and the accepted true value, together with confidence intervals.

For accuracy testing, 10mg of Ciprofloxacin HCL standard was weighed in an electronic balance and then transferred it to a 100 ml volumetric flask and it was then dissolved properly with distilled water. The solution finally was made up to the mark with distilled water. 10 ml of solution was then transferred to another 100ml volumetric flask to dilute the solution 10 times with distilled water. Four replicates were prepared by the same way and the absorbance values were taken at 272 nm.

2.2.4.4 Recovery Testing

While making the calibration curve, three replicates of each concentration was prepared and the average absorbance values were calculated. Calibration curve was prepared by constructing the concentration vs. average absorbance values. Again from the average

absorbance values, the observed concentration was calculated using the equation of calibration curve. The percentage of recovery was then calculated.

2.2.4.5 Precision

The precision of an analytical procedure is the degree of agreement between individual test results when the procedure is applied repeatedly to multiple samplings or a homogeneous sample. The precision of an analytical procedure is usually expressed as the standard deviation or relative standard deviation (coefficient of variation) of a series of measurements. Precision may be a measure of either the degree of reproducibility or of repeatability of the analytical procedure under normal operating conditions. In this context, reproducibility refers to the use of the analytical procedure in different laboratories, as in a collaborative study. Intermediate precision (also known as ruggedness) expresses within-laboratory variation, as on different days, or with different analysts or equipment within the same laboratory. Repeatability refers to the use of the analytical procedure within a laboratory over a short period of time using the same analyst with same equipment. Three different concentrations within the range highest, middle and the lowest were prepared three times. The solutions were

analyzed by its absorbance values keeping them at room temperature two times in a day (intraday precision) and at 1 and 4 days intervals (inter day precision). The values were compared with the initial absorbance values in terms of RSD [8].

2.2.4.6 Sensitivity

Limit of detection (LOD)

The detection limit is a characteristic of limit tests. It is the lowest amount of analyte in a sample that can be detected, but not necessarily quantitated, under the stated experimental conditions. Thus, limit tests merely substantiate that the amount of analyte is above or below a certain level. The detection limit is usually expressed as the concentration of analyte (e.g., percentage, parts per billion) in the sample".

In order to find out LOD, concentrations of Ciprofloxacin HCL in distilled water from 0.1 to 2 μ g/mL was prepared and absorbance values were recorded. The concentration which showed the minimum absorbance value was selected as the LOD of this UV method.

2.2.4.7 Assay of Ciprofloxacin HCL

Three tablets of two strengths (250 and 500 mg) of Ciprofloxacin HCL marketed brand product were purchased and average weights were noted. The three tablets were grinded using mortar and pestle and a definite amount of the tablet powder was transferred in a 100 mL, volumetric flask. Distilled water

and methanol at 90:10 ratio was added to dissolve the powder mix. The mixture was then sonicated for 15 mins and then made upto volume. The solution was finally filtered using whatman filter paper. The clear solution was again diluted using the same water methanol (90:10) mixture and absorbance values were recorded.

2.2.5 Statistical analysis

All the data were analyzed using MS Excel 2007. Most of the cases the results were represented as mean \pm standard deviation (SD).

Limit of Quantitation (LOQ)

The quantitation limit is a characteristic of quantitative assays for low levels of compounds in sample matrices, such as impurities in bulk drug substances and degradation products in finished pharmaceuticals. It is the lowest amount of analyte in a sample that can be determined with acceptable precision and accuracy under the stated experimental conditions. The quantitation limit is expressed as the concentration of analyte (e.g., percentage parts per billion) in the sample".

In order to find out LOQ, concentrations of Ciprofloxacin HCL in distilled water from 0.1 to 2 μ g/mL was prepared and absorbance values were recorded. The concentration which showed the minimum absorbance value with acceptable recovery and precision values

was selected as the LOQ of this UV method.

RESULTS AND DISCUSSION

3.1 Construction of calibration curve

To construct the calibration curve seven standard solutions were prepared (2, 3, 4, 6, 8, 10, 12 µg/ml concentration). The absorbance values of these solutions were noted at 272nm wavelength (λ_{\max} for Ciprofloxacin HCL). Data was collected and plotted (**Table 4**) in a chart to obtain the standard curve. From this curve a regression coefficient value (R^2) of 0.9997 was found.

By plotting the absorbance against the concentration of Ciprofloxacin a straight line was found. The equation obtained from the calibration curve was $Y=0.0826x + 0.007$.

3.2 Method validation parameter

3.2.1 Linearity

The linearity was evaluated by analyzing three different calibration curves of Ciprofloxacin HCL. The regression coefficient and slope were specifically evaluated by using the RSD value (**Table 5**).

3.2.2 Range

The method obeyed the Beer-Lambert's law from the concentration range of 2-12 µg/mL.

3.2.3 Accuracy testing result

To study the accuracy of the proposed methods, and to check the interference from excipients used in the dosage forms,

recovery experiments were carried out by the standard addition method. Accuracy testing data is shown in **Table 6**. The table shown that $103.36 \pm 7.93\%$ Ciprofloxacin was obtained from the assay testing. $100 \pm 5\%$ is an acceptable range. Therefore, the value was slightly higher than the recommended range.

3.2.4 Recovery testing

The accuracy of the method was determined by recovery experiments. A known amount of standard Ciprofloxacin hydrochloride corresponding to 2, 3, 4, 6 and 8, 10, 12% of the label claim (standard addition method) was added to pre-analyzed sample of tablet. The acceptable value of recovery is 80-110%. From the recovery data **Table 7** we can see that the mean recovery was 100.69%. Therefore, the value was in recommended range.

3.2.5 Intra and Inter day precision

In this step of method validation we examined the differences in UV absorbance of Ciprofloxacin standard solutions at different time in a particular day and at different days. To do so 3 sets of 2, 6 and 12 µg/mL solutions were prepared from 100 µg/mL stock solution and the solutions were analyzed in the morning and afternoon and then after 2 days and after 4 days respectively.

Intraday precision result:

The concentrations of the drug were measured three times on the same day at intervals of five hours. From the table 7 we can see the average concentration of the drug taken in the morning and the evening was almost same. The average conc. in the morning for 2, 6, and 12 was 2.21, 6.13 and 12.32 respectively and the average conc. in the evening for 2,6,12 was 2.39, 6.32, and 12.58 respectively. A negligible variation was observed. The acceptable RSD value is 5. We got the RSD value 3.13 which is in recommended range. The RSD value was given in the **Table 8**.

Inter day precision result: One day interval:

The concentrations of the drug was measured on three different days for inter day study. The standard deviation and Relative Standard Deviation were calculated (RSD). Precision data of one day interval is shown in the table 8. The average concentration even after one day interval remained almost same. The average concentration in day zero for 2, 6 and 12 was 3.74, 6.51 and 11.19 respectively and the average concentration after one day for 2, 6 and 12 was 3.94, 6.83 and 11.36 respectively. The acceptable RSD value is 5. We got the RSD value 2.75 which is in recommended range. The RSD value was given in the **Table 9**.

Four days interval:

The concentrations of the drug was measured on three different days for inter day study. The standard deviation and Relative Standard Deviation were calculated (RSD). The average concentration of drug after four days interval remained almost same except the concentration 12. The concentration of 12 after four days was notably changed compared to concentration 2 and 6. The average concentration in day zero for 2, 6 and 12 was 3.74, 6.51 and 11.19 respectively and the average concentration after four days for 2, 6 and 12 was 3.96, 6.87 and 10.21 respectively. It is also noticed that the concentration of 2 and 6 increased slightly whereas the concentration of 12 decreased slightly. The acceptable RSD value is 5. We got the RSD value 4.73 which is in recommended range.

3.2.6 Sensitivity of the method

The LOD of Ciprofloxacin was found to be 0.1 µg/ml and LOQ was found to be 2 µg/ml.

3.2.7 Assay testing of 250 and 500 mg of Ciprofloxacin HCl tablet

To perform the assay testing of Ciprofloxacin HCl tablet of a renowned pharmaceutical company of India was used as the known standard. Each of this ciprocin[®] tablet contains 250 and 500 mg of Ciprofloxacin HCl. 3 tablets from each strength were weighed in an electronic balance and the average weight was calculated.

The recommended value of % recovery is 80-110%. From the **Table 10** we can see the obtained % recovery of 250 mg of Ciprofloxacin HCL is 87.10 %. Which is within the recommended range. The

recommended value of % recovery is 80-110%. From the **Table 11** we can see the obtained % recovery of 500 mg of Ciprofloxacin HCL is 88.55 %. Which is within the recommended range.

Table 4: Absorbance data of standard solutions of Ciprofloxacin at 272 nm

Conc. ($\mu\text{g/mL}$)	Standard 1	Standard 2	Standard 3	Average
2	0.171	0.167	0.178	0.172
3	0.255	0.253	0.258	0.255
4	0.336	0.334	0.339	0.336
6	0.501	0.508	0.5	0.503
8	0.663	0.658	0.664	0.662
10	0.839	0.849	0.844	0.844
12	1.013	0.958	1.005	0.992

Table 5: R^2 , slope and intercept values of Ciprofloxacin HCL in three standard solutions

Conc. ($\mu\text{g/ml}$)	Standard 1	Standard 2	Standard 3	Average	SD	RSD
R^2	0.999888	0.998574	0.999932	0.999465	0.0008	0.08
Slope (m)	0.083855	0.080843	0.08298	0.082559	0.0015	1.88
Intercept	0.000647	0.012724	0.007703	0.007024	0.0061	0.07

Table 6: Accuracy data of Ciprofloxacin HCl

Absorbance	Dilution factor	Observed conc.	Original conc.	% Recovery	Mean	SD
0.938	10	107.87	100.00	107.87	103.36	7.93
0.935	10	112.35	100.00	112.35	102.33	6.29
0.885	10	106.30	100.00	106.30	103.22	5.39
0.778	10	93.34	100.00	93.34	101.33	8.32
0.137	10	98.57	100.00	96.93	100.45	6.38

Table 7: Recovery data of Ciprofloxacin HCl

Conc. ($\mu\text{g/mL}$)	Standard 1	Standard 2	Standard 3	Average	Observed conc. ($\mu\text{g/mL}$)	%Recovery
2	0.171	0.167	0.178	0.172	2.01	100.61
3	0.255	0.253	0.258	0.255	3.03	100.95
4	0.336	0.334	0.339	0.336	4.02	100.41
6	0.501	0.508	0.5	0.503	6.05	100.81
8	0.663	0.658	0.664	0.662	7.98	99.80
10	0.839	0.849	0.844	0.844	10.21	102.07
12	1.013	0.958	1.005	0.992	12.01	100.10
Mean Recovery%						100.69
SD Recovery						0.24
RSD recovery						0.24

Table 8: Intraday precision data of Ciprofloxacin HCL

Conc. ($\mu\text{g}/\text{mL}$)	Abs. (Day 0)	Abs. (Day 2)	Conc. (Day 0)	Conc. (Day 2)	Avg. conc. in Day 0	Avg. conc. in Day 2	Mean conc. (bet. Day 0 & Day 2)	SD of conc. (bet. Day 0 & Day 2)	RSD	Avg. RSD
2	0.248	0.247	2.94	3.10	3.74	3.94	3.84	0.14	3.74	2.75
	0.313	0.317	3.73	3.95						
	0.38	0.385	4.55	4.78						
6	0.566	0.574	6.82	7.09	6.51	6.83	6.67	0.23	3.45	
	0.538	0.553	6.48	6.83						
	0.518	0.533	6.23	6.59						
12	0.912	0.912	11.04	11.21	11.19	11.36	11.27	0.12	1.07	
	0.927	0.927	11.22	11.39						
	0.934	0.934	11.30	11.48						

Table 9: Inter day (1 day) precision data of Ciprofloxacin HCl

Conc. ($\mu\text{g}/\text{mL}$)	Abs. (obtd. in the morn.)	Abs. (obtd. in the eve.)	Conc. (obtd. in the morn.)	Conc. (obtd. in the eve.)	Avg. conc. in the morn.)	Avg. conc. in the eve.)	Mean conc. (bet. mom. eve.)	SD of conc. (bet. morn. eve.)	RSD	Avg. RSD
2	0.188	0.19	2.21	2.40	2.21	2.39	2.30	0.13	5.75	3.13
	0.186	0.187	2.18	2.37						
	0.19	0.191	2.23	2.41						
6	0.51	0.51	6.13	6.30	6.13	6.32	6.23	0.14	2.17	5.28
	0.51	0.514	6.13	6.35						
	0.509	0.51	6.12	6.30						
12	1.025	1.031	12.41	12.66	12.32	12.58	12.45	0.18	1.48	10.12
	1.016	1.023	12.30	12.56						
	1.01	1.019	12.23	12.51						

Table 10: %Recovery of 250 mg of Ciprofloxacin HCL tablet

Dilution factor	Absorbance	Concentration ($\mu\text{g}/\text{ml}$)	Known concentration ($\mu\text{g}/\text{mL}$)	% Recovery
50	0.467	279.06	319.13	87.10

Table 11: %Recovery of 500 mg of Ciprofloxacin HCL tablet

Dilution factor	Absorbance	Concentration ($\mu\text{g}/\text{ml}$)	Known concentration ($\mu\text{g}/\text{mL}$)	% Recovery
50	0.468	279.06	315.13	88.55

3.2.9 DISCUSSION

Estimation of Ciprofloxacin Hydrochloride was found to be simple, accurate and reproducible. It follows Beer-Lambert's law in the concentration range of 2-12 $\mu\text{g}/\text{ml}$. The optical characteristics such

as percent relative standard deviation and percent range of error were found to be within the limit and satisfactory. All of the analytical validation parameter for the proposed method was determined according to USP guidelines. All

validation parameters results were getting within the range of USP standards.

The recovery studies showed that the result were within the limit indicating no interference. The proposed method is simple, sensitive, accurate and precise and can be successfully employed for the routine analysis of the Ciprofloxacin hydrochloride in bulk drug and solid dosage forms.

3.2.10 CONCLUSION

The statistical analyses showed that the data from the proposed method are in good agreement for the estimation of Ciprofloxacin hydrochloride in bulk drug and solid dosage forms. The method is economical, rapid and do not require any sophisticated instruments contrast to chromatographic method. Thus it can be effectively applied for the routine analysis of ciprofloxacin hydrochloride in bulk drug and solid dosage forms.

REFERENCES

- [1] Olaniyi, A.A., Babalola, Oladeinde, F. and Adegoke, A., 2001. Towards better quality assurance of drugs. In Proceedings of the 4th National Workshop pp. 70-73.
- [2] Adegbolagun, O.M., Jegede, O.J. and Olaniyi, A.A., 2002. Chemical Equivalence studies on four brands of metronidazole tablets. *Nig. J. Pharm. Res.* 2002; (April-Sept): 71, 74.
- [3] Adegbolagun, O.A., Olalade, O.A. and Osumah, S.E., 2007. Comparative evaluation of the biopharmaceutical and chemical equivalence of some

commercially available brands of ciprofloxacin hydrochloride tablets. *Tropical journal of pharmaceutical research*, 6(3), pp.737-745

- [4] Nijhu. R.S., Jhanker, Y.M. and Sutradhar, K.B., 2011. Development of an assay method for simultaneous determination of ciprofloxacin and naproxen by UV spectrophotometric method. *Stamford journal of pharmaceutical sciences*, 4(1), pp. 84- 90.
- [5] Slama TG. Serum sickness-like illness associated with ciprofloxacin. *Antimicrob Agents Chemother* 1990; 34: 904-5
- [6] Wolfson JS, Hooper DC. Overview of fluoroquinolone safety. *Am J MED* 1991; 91.
- [7] The United States of Pharmacopoeia (2011), pp. 776-778.
- [8] Varanda, F., Pratas de-Melo, M.J., Caw, A.I., Dohrn, R., Makrydaki, F.A., Voutsas, E, Tassios, D. and Marrucho, M., 2006. Solubility of antibiotics in different solvents. I Hydrochloride forms of tetracycline, moxifloxacin, and ciprofloxacin. *Industrial & engineering chemistry research*. 45(18). pp.6368-63.